

Test Report

SL52025257636301TX

Date: June 12, 2020

Page 1 of 3

HENAN GORE MEDICAL INSTRUMENTS CO., LTD
THE NORTH INDUSTRIAL AREA OF DING LUAN, CHANGYUAN COUNTRY, 453412 HENAN, CHINA

The following sample(s) was/were submitted and identified on behalf of the client as:

Sample Description : (A) Medical protective mask (Claimed Type II)

Style No. : arch

Sample Color : White

Sample Receiving Date : May 15, 2020

Testing Period : May 22, 2020 - Jun 12, 2020

Test Result(s) : Unless otherwise stated the results shown in this test report refer only to the sample(s) tested, for further details, please refer to the following page(s).

Test Performed : Selected test(s) as requested by applicant

Comment:

Medical Face Masks-Requirements and Test Methods(EN 14683:2019+AC:2019)	(A)
Clause 5.2.2 Bacterial filtration efficiency (BFE)	M
Clause 5.2.3 Breathability	M
Clause 5.2.5 Microbial Cleanliness	M

Remark: M=Meet EN 14683:2019+AC:2019 Type II requirement

Signed for and on behalf of
SGS-CSTC Standards Technical Services (Shanghai) Co., Ltd Testing Center

Sara Guo (Account Executive)

Dongjing Liu (Authorized Signatory)



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

3rd Building, No. 889, Yishan Road, Xuhui District Shanghai, China 200233 t (86-21) 61402666 f (86-21) 64958763 www.sgsgroup.com.cn
中国·上海·徐汇区宜山路889号3号楼 邮编: 200233 t (86-21) 61402666 f (86-21) 64958763 e sgs.china@sgs.com

Test Result

Medical Face Masks-Requirements and Test Methods
(EN 14683:2019+AC:2019)

Clause 5.2.2 Bacterial filtration efficiency (BFE)
(EN 14683:2019 Annex B)

Sample: A

(BFE), %	1#	2#	3#	4#	5#
	99.9	99.7	99.8	99.7	99.9

Remark: Performance Requirement: Type I $\geq 95\%$, Type II $\geq 98\%$, Type IIR $\geq 98\%$
* This test standard is not within the accredited scope in SGS Shanghai testing centre, it is carried out by external laboratory accredited by CMA (China Metrology Accreditation).

Clause 5.2.3 Breathability (Differential Pressure)
(EN 14683:2019 Annex C, Flow rate 8 l/min)

Sample: A

Differential pressure ΔP (Pa/cm ²)	1#	2#	3#	4#	5#
	26	26	28	28	26

Remark: Performance Requirement: Type I < 40 Pa/cm², Type II < 40 Pa/cm², Type IIR < 60 Pa/cm²

Clause 5.2.5 Microbial Cleanliness
(EN 14683:2019+AC:2019 Annex D and EN ISO 11737-1:2018)

Sample: A

CFU/g	1#	2#	3#	4#	5#
	1	<1	<1	<1	<1

Remark: Performance Requirement: Type I ≤ 30 CFU/g, Type II ≤ 30 CFU/g, Type IIR ≤ 30 CFU/g

Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/sgs/terms-and-conditions.aspx>, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/sgs/terms-and-conditions.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained herein reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the samples tested and such samples are retained for 30 days only.
Attention: To check the authenticity of testing inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CH.Overseas@sgs.com
Building No.888, Yehou Road, Xuhui District, Shanghai, China 200233 | T: (86-21) 61402666 | F: (86-21) 61466793 | www.sgs.com | www.sgs.com.cn
中国·上海·徐汇区宜山路888号3号楼 邮编: 200233 | T: (86-21) 61402666 | F: (86-21) 61466793 | e: sgs.china@sgs.com

Member of the SGS Group (SGS SA)

Sample Photo



The statement of conformity in this test report is only based on measured values by the laboratory and does not take their uncertainties into consideration.

End of Report

Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/sgs/terms-and-conditions.aspx>, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/sgs/terms-and-conditions.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained herein reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the samples tested and such samples are retained for 30 days only.
Attention: To check the authenticity of testing inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CH.Overseas@sgs.com
Building No.888, Yehou Road, Xuhui District, Shanghai, China 200233 | T: (86-21) 61402666 | F: (86-21) 61466793 | www.sgs.com | www.sgs.com.cn
中国·上海·徐汇区宜山路888号3号楼 邮编: 200233 | T: (86-21) 61402666 | F: (86-21) 61466793 | e: sgs.china@sgs.com

Member of the SGS Group (SGS SA)

产品说明书

- CE Mark: conforms to essential requirements of the Medical Device Regulation 2017/745 (EU)
 - Date of manufacture. Indicates the date when the medical device was manufactured
 - Use-by date. Indicates the date after which the medical device is not to be used
 - Do not re-use, indicates a medical device that is intended for one use, or for use on a single patient during a single procedure
 - Do not use if package is damaged, indicates a medical device that should not be used if the package has been damaged or opened
 - Non-sterile
 - UDI-DI code of product identification
 - Batch code. Indicates the manufacturer's batch code so that the batch or lot can be identified.
 - Fragile, handle with care, indicates a medical device that can be broken or damaged if not handled carefully.
 - Keep away from sunlight, indicates a medical device that needs protection from light sources.
 - Keep dry, indicates a medical device that needs to be protected from moisture.
- Name: Henan Gore Medical Instruments Co., Ltd.
Add: The North Industrial Area of Ding Looan, Changyuan County, 453412 Henan, China.
Tel: +86-0373-8969880
Fax: +86-0373-8967288
- Name: Llin Service & Consulting GmbH
Address: Obere Seegasse 34/2, 69124, Heidelberg, Germany
DIMDI Code: DE:0000048234
Tel: +49 175 4870 819
Email: info@llin-service.com

产品说明书



- Medical protective mask
- Instruction for Use
- [Product Name]**
Medical protective mask
- [Model/Size]**
Rectangle, Arch, Butterfly
- [Intended use]**
It is used for the protection of medical units in non-invasive operation.
- [Contraindications]**
People who are allergic to product materials.
- [Directions for use]**
1. Open the package, take out the mask and expand it longitudinally to form a space between the nose and mouth to make the breathing smooth.
 2. Face the mask with light color, dark color and nose clip upward (a plastic strip).
 3. After wearing the face, adjust the nose clip to make it have good tightness and prevent harmful substances in the air from leaking into the sealing place.
 4. After use, it shall be discarded in the specified dirt box and treated in a unified way to avoid product reuse hazards and environmental pollution.
- [Precautions]**
1. Read the manual carefully before use;
 2. This product is disposable, can not be washed and reused;
 3. Do not use in sterile ward or invasive operating room;
 4. Pay attention to the "non sterile" and "sterile" instruction marks of the product packaging, please use them after handling according to the clinical requirements;
 5. It is forbidden to use the package if the seal is disconnected, damaged or beyond the validity period of the product.
- [Package]**
Plastic packaging bag, Single packing.
- [Expiry]**
2 years.
- [Production Date]**
Refer to the package label
- [Storage conditions]**
1. Products shall be protected from heavy pressure, direct sunlight and rain and snow during transportation;
 2. It shall be stored in a ventilated, dry and non corrosive gas environment. Keep away from fire sources and inflammables.
- [Sterilization Method]**
None
- [Symbol Description]**
- Caution, Indicates the need for the user to consult the instructions for use for important cautionary information, such as warnings and precautions that cannot, for a variety of reasons, be presented on the medical device itself.



CMC MEDICAL DEVICES & DRUGS SL
NO. CMC/CE/2020/04062020.15

CONFIRMED THAT CMC MEDICAL DEVICES & DRUGS S.L. Is the European Authorized Representative of

Henan Gore Medical Instruments Co., Ltd.
The North Industrial Area of Ding Luan, Changyuan County, 453412 Henan, China

The certificate remains valid until the expiration agreement of EC REP, manufacturing conditions, the quality system or relevant legislation are changed. The validity is conditioned by positive results of periodic surveillance audits.
The product liability rests with the manufacturer in accordance with applicable directive and standard, after fulfilling of the relevant EU legislation requirements, the manufacturer shall affix relevant CE marking to all above mentioned models of the medical device.

Complies with the applicable essential requirements of the council directive 93/42/EEC on medical devices as amended.

The products in Annex I was registered in Spanish MOH with number **RPS/1228/2020**



Issued on: 04/06/2020



Authorized Signatory
CMC Medical Devices & Drugs SL

Valid until: 03/06/2021



ANNEX I Medical Device Products

Protective clothing

Surgical mask

Disposable medical gloves

Medical protective mask

Disposable isolation clothing



Certificate Of Registration

HENAN GORE MEDICAL INSTRUMENTS CO., LTD
The North Industrial Area of Ding Luan, Changyuan County, 453412 Henan, China

Has Completed With The U.S. Food And Drug Administration Pursuant To 21 CFR Part 807: Establishment Registration And Device Listing

Owner/Operator No.: 10065726

Device Listing: See Next pager

Huawin will confirm that such registration remains effective upon request and presentation of this certificate until the end of the calendar year stated above, unless said registration is terminated after issuance of this certificate. Huawin makes no other representations or warranties, nor does this certificate make sole benefit it is issued. This certificate does not denote endorsement or approval of the certificate-holder's device or establishment by the U.S. Food and Drug Administration. Huawin assumes no liability to any person or entity in connection with the foregoing.

The U.S. Food and Drug Administration does not issue a certificate of registration, nor does the U.S. Food and Drug Administration recognize a certificate of registration. Huawin is not affiliated with the U.S. Food and Drug Administration.



Manager: Guy Su
Issue date: May 28, 2020
Expire Date: Dec. 31, 2020



Certificate Of Registration

Device Listing:

Listing No.	Code No.	Proprietary Name	Model No.
D382100	KHA	Surgical mask(Ear loop/Tie band)	-
		Disposable Medical Mask(Ear loop/Tie band)	-
		Medical protective mask	-
D382098	OEA	Disposable isolation clothing(one-piece type)	-
		Disposable medical protective clothing(split type/one-piece type)	-
		Disposable isolation clothing(split type)	-



Manager: Guy Su
Issue date: May 28, 2020
Expire Date: Dec. 31, 2020



Product Service

Certificate

No. Q6 104197 0001 Rev. 00

Holder of Certificate: **Henan Gore Medical Instruments Co., Ltd.**

The North Industrial Area of Ding Luan
453412 Changyuan County, Henan
PEOPLE'S REPUBLIC OF CHINA

Facility(ies):

Henan Gore Medical Instruments Co., Ltd.
The North Industrial Area of Ding Luan, 453412 Changyuan
County, Henan, PEOPLE'S REPUBLIC OF CHINA

Certification Mark:



Scope of Certificate:

Production and Distribution of Disposable Medical Mask, Disposable Medical Cap, Medical Absorbent Gauze Patches, Medical Cotton Swab, Disposable Medical Bedsheet, Disposable Surgical Gown, Disposable Anesthesia Gas Filter, Medical Oxygen Mask, Disposable Airflow Atomizer, Trachea Cannula, Medical Bandage, Disposable Intravenous Indwelling Needle

Applied Standard(s):

EN ISO 13485:2016
Medical devices - Quality management systems -
Requirements for regulatory purposes
(ISO 13485:2016)
DIN EN ISO 13485:2016

The Certification Body of TÜV SÜD Product Service GmbH certifies that the company mentioned above has established and is maintaining a quality management system (excluding subclause 7.3), which meets the requirements of the listed standard(s). See also notes overleaf.

Report No.: SH19149901
Valid from: 2020-03-11
Valid until: 2023-03-10

Date, 2020-03-11

Christoph Dicks
Head of Certification/Notified Body

7/15

Benannt durch/Designated by
Zentralstelle der Länder
für Gesundheitsschutz
bei Arzneimitteln und
Medizinprodukten
www.zlg.de
ZLG-BS-244.10.08



Product Service

EC Certificate

Production Quality Assurance System

Directive 93/42/EEC on Medical Devices (MDD), Annex V

(Devices in class I in sterile conditions, sterilised systems or procedure packs)

No. G2S 104197 0003 Rev. 00

Manufacturer

Henan Gore Medical Instruments Co., Ltd.

The North Industrial Area of Ding Luan
453412 Changyuan County, Henan
PEOPLE'S REPUBLIC OF CHINA

Product Category(ies):

**Disposable Medical Mask,
Disposable Medical Cap,
Medical Absorbent Gauze Patches,
Medical Cotton Swab,
Disposable Medical Bedsheet,
Disposable Surgical Gown**

The Certification Body of TÜV SÜD Product Service GmbH declares that the aforementioned manufacturer has implemented a quality assurance system for manufacture in accordance with MDD Annex V. This quality assurance system covers those aspects of manufacture concerned with securing and maintaining sterile conditions of the respective devices / device categories and conforms to the requirements of this Directive. It is subject to periodical surveillance. See also notes overleaf.

Report No.: SH19149901

Valid from: 2020-03-11

Valid until: 2024-05-26

Date, 2020-03-11

Christoph Dicks
Head of Certification/Notified Body

南德 TUV 认证产品清单

S/N	Category	Name of Product	Types/Sizes/ Models	Classification/ Rule of classification	UMDNS Code
1	MD 0101 MDS7006_1	Trachea Cannula(气管插管)	Ordinary Type: 2.0-11.0 Reinforced Type: 2.0-11.0	Class IIa/ Rule 5	14085
2	MD 0101 MDS7006_1	Disposable Anaesthesia Gas Filter (一次性使用麻醉气体过过滤器)	Normal (常规)	Class IIa/ Rule 2	15649
3	MD 0101 MDS7006_1	Medical Oxygen Mask (医用吸氧面罩)	XL, L, M, S	Class IIa/ Rule 2	12448
4	MD 0101 MDS7006_1	Disposable Airflow Atomizer (一次性使用气流雾化器)	Type I (大杯): mouthpiece, mask(XL/L/M/S), Type II (小杯): mouthpiece, mask(XL/L/M/S)	Class IIa/ Rule 2	10226
5	MD 0301 MDS7006_1	Disposable Medical Cap (一次性使用医用帽)	Straight (直筒型), cylinder (圆顶形), strip (条形)	Class Is/ rule 1	16081
6	MD 0301 MDS7006_1	Disposable Medical Mask (一次性使用医用口罩)	Lacing (系带型), hanger (挂耳型)	Class Is/ rule 1	12447
7	MD 0301 MDS7006_1	Medical Absorbent Gauze Patches (医用脱脂纱布块)	Length (2cm~1000cm) × Width(2cm~80cm) × Layers(1~32)	Class Is/ rule 4	11859
8	MD 0301 MDS7006_1	Medical Cotton Swab (医用棉拭)	6cm, 8cm, 10cm, 12cm, 15cm, 16cm, 20cm, 22cm, 25cm	Class Is/ rule 4	15066
9	MD 0301 MDS7006_1	Disposable Medical Bedsheet (一次性使用医用床单)	Type I, Type II, Type III	Class Is/ rule 1	15075
10	MD 0301 MDS7006_1	Disposable Surgical Gown (一次性使用手术衣)	extra large, large, middle, small, extra small (特大、大号、中号、小号)	Class Is/ rule 1	11901

CE 一类备案清单

一次性医用防护服 protective clothing	型号: 连身式、分身式; 规格: 160、165、170、175、180、185、190 Model: one-piece, two-piece; Specification: 160, 165, 170, 175, 180, 185, 190
医用外科口罩 Surgical mask	系带型、挂耳型 lacing, hanger
一次性使用医用橡胶检查手套 Disposable medical rubber inspection gloves	5、5.5、6、6.5、7、7.5、8、8.5、9、9.5
医用防护口罩 Medical protective mask	蝶型、拱形、长方形 Butterfly, arch, rectangle
一次性使用隔离衣 Disposable isolation clothing	型号: 连身式、分身式、褂式; 规格: 160、165、170、180、185、190; 褂式: 110、115、120、130、140、150、160、170、180、190 Model: one-piece, two-piece, jacket; Specification: 160, 165, 170, 180, 185; Jacket style: 110, 115, 120, 130, 140, 150, 160, 170, 180, 190

河南省医疗器械检验所

Henan Province Medical Instrument Testing Institute

检验检测报告

Inspection Report

Product Name: Medical Protective Mask

Inspection Type: Emergency Inspection

Client: Henan Drug Administration



Henan Province Medical Instrument Testing Institute

Inspection Report

Report No: 202001513

Sample No: E20200354


Page 3 of 1

sample name	Medical protective mask	Sample Q'ty	50pcs
	Sample sending () sampling (✓)	Specification model	Arched L
Client	Henan Drug Administration	Production batch	20022101
Address	North Industrial Zone, Dingluan Town, Changyuan County	Production Date	21 Feb., 2020
Production company	Henan Gore Medical Instruments Co., Ltd.	Product No	/
Tested Company	Henan Gore Medical Instruments Co., Ltd.	Expiry date	20 Feb., 2022
Sampling Company	/	Inspection type	Emergency inspection
Quantity in stock	/	Sample status	normal
Sampling date	/	Sample collection date	2020.03.17
Sampling location	/	Inspection location	Laboratory of this inspection institute
Sample number	/	Inspection date	2020.03.17- 2020.03.25
Test items	Some items		
Test Standard	GB 19083-2010 《Technical requirements for medical protective masks》		
Test results	The inspected items meet requirements of GB 19083-2010《Technical Requirements for Medical Protective Masks》 Date of issue: March 26, 2020		
Remarks	In the report, "----"Indicates that this item is not applicable; In the report, " / "Indicates that this item is blank or non-test.		


Approved by:



Checked by:



Inspector:





Henan Province Medical Instrument Testing Institute

Inspection Report

Sample No: E20200354

Page 3 of 2

Medical protective mask				
Test items	Standard Terms	standard requirement	test result	conclusion
Nose clip	4.2.1	Nose clip should be on the mask	Meet requirements	qualified
	4.2.2	Nose clips should be adjustable	Meet requirements	qualified
Mask belt	4.3.1	Mask belt should be easy to adjust	Meet requirements	qualified
	4.3.2	There should be sufficient strength to fix the position of the mask. The breaking strength of the connection point between mask belt and mask body should not be less than 10N.	Meet requirements	qualified
Filtration efficiency	4.4	At flow rate of 85 L / min, non-oily particles filtration efficiency should meet requirements of Table 1. (95%)	Before temperature pretreatment Mini: 97%, after temperature pretreatment mini: 98%	qualified
Synthetic blood penetration	4.6	Spray 2mL of Synthetic blood at pressure 10.7kPa (80mmHg) to the mask, should be no penetration to inner side of mask.	Meet the requirements	qualified
Microbiological index	4.8.1	Masks should meet the requirements of microbiological indicators in GB15979-2002, see Table 2	Bacteria < 20cfu/g, fungus < 20 / g, coliform bacteria, Staphylococcus aureus, P. aeruginosa Bacteria and hemolytic streptococci were not detected.	qualified
Leakproofness	4.12	The mask design should provide good leakproofness, and the total fit factors of the mask should be no less than 100	Meet requirements	qualified
<p>Note: Blank</p>				

仪器



中国认可
国际互认
检测
TESTING
CNAS L10118



国检检测
CHINA COMPONENTS TEST

Test Report

(2020) WSZ FHL NO.W381

Product Name Filtering half mask

Client Henan Gore Medical Instruments Co., Ltd.

Manufacturer Henan Gore Medical Instruments Co., Ltd.

Test Type Entrusted inspection


Jiangsu Guojian Testing Technology Co., Ltd.

检验专用章

Test Report

[2020] WSZ FHL NO.5420

Page 1 of 5

Product name	Filtering half mask	Specification	—
		Brand	—
Client/Add/Tel	Henan Gore Medical Instruments Co., Ltd/The North Industrial Area of Ding Luan, Changyuan County, 453412 Henan, China		
Manufacturer/Add/Tel	Henan Gore Medical Instruments Co., Ltd/The North Industrial Area of Ding Luan, Changyuan County, 453412 Henan, China		
Sample grade	FFP2	Sample number	GW0381-2020
Sample quantity	110 pcs	Receiving date of sample	11/05/2020
Test type	Entrusted inspection	Article number/Batch number/Style number	—
Test date	16/05/2020~03/06/2020	Testing sites	Testing room
Sample state	Meeting the requirements of testing	Sample description	—
Test standard(s)	EN 149:2001+A1:2009 Respiratory protective devices-Filtering half masks to protect against particles- Requirements,testing, marking		
Test items	Visual inspection, practical performance, finish of parts, compatibility with skin, flammability, carbon dioxide content of the inhalation air, material, head harness, field of vision, penetration of filter material, breathing resistance, total inward leakage		
Test conclusion	<p>The sample upon testing, the test items meet the requirements of the EN 149:2001+A1:2009 standard. The detail of test results see on Pages 2-5.</p> <p style="text-align: right;">Issue date: 16/06/2020</p>		
Note	For the entrusted sample test, the technical responsibilities are undertaken for the test results of the supplied samples only.		

Approver:

苏晓群

Reviewer:

王明

Chief Tester:

杨莹

Test Report

[2020] WSZ FHL NO.5420

Page 2 of 5

S.No	Test item	Unit	Technical requirements	Test result	Single item decision
1	Visual inspection	Packaging	Particle filtering half masks shall be offered for sale packaged in such a way that they are protected against mechanical damage and contamination before use.	Packaging withstands mechanical damage and contamination.	Qualified
		Material	Materials used shall be suitable to withstand handling and wear over the period for which the particle filtering half mask is designed to be used.	Materials withstand handling and wear.	
2	Practical performance	Head harness comfort	Head harness should be comfort.	Sample 1 has the feeling of comfortable wearing Sample 2 has the feeling of comfortable wearing	Qualified
		Security of fastenings	Fastenings are safe and reliable	Sample 1: All fastenings are firm. Sample 2: All fastenings are firm	
		Field of vision	Field of vision is acceptable	Sample 1: Having a wider visual field Sample 2: Having a wider visual field	
3	Finish of parts	—	Parts of the device likely to come into contact with the wearer shall have no sharp edges or burrs.	Parts of the device have no sharp edges and burrs	Qualified
4	Compatibility with skin	—	Materials that may come into contact with the wearer's skin shall not be known to be likely to cause irritation or any other adverse effect to health.	A.R. 5 pcs all don't cause irritation T.C. 5 pcs all don't cause irritation	Qualified
		—	When tested, the particle filtering half mask shall not burn or not to continue to burn for more than 5s after removal from the flame.	A.R. The Sample is burning. Burning time:0.1s The Sample is burning. Burning time:0.1s T.C. The Sample is burning. Burning time:0.1s The Sample is burning. Burning time:0.1s	

Test Report

[2020] WSZ FHL NO.5420

Page 3 of 5

S.No	Test item	Unit	Technical requirements	Test result			Single item decision	
				Sample 1	Sample 2	Sample 3		
6	Carbon dioxide content of the inhalation air	—	$\leq 1.0\%$ (by volume)	0.7110%	0.7120%	0.7130%	Qualified	
				Report value	0.71%			
				Sample 1: neither facepiece nor straps have mechanical failure				
				Sample 2: neither facepiece nor straps have mechanical failure				
7	Material	—	After undergoing S.W., none of the particle filtering half masks shall have suffered mechanical failure of the facepiece or straps. After undergoing S.W. and T.C., none of the particle filtering half masks shall not collapse.	Sample 1: neither facepiece nor straps have mechanical failure			Qualified	
				Sample 2: neither facepiece nor straps have mechanical failure				
				Sample 3: neither facepiece nor straps have mechanical failure				
				Sample 1: no collapse				
8	Head harness	—	The head harness shall be designed so that the particle filtering half mask can be donned and removed easily The head harness shall be adjustable or self-adjusting and shall be sufficiently robust to hold the particle filtering half mask firmly in position	A.R.	All of 5 pieces particle filtering half mask meet the requirements		Qualified	
				T.C.	All of 5 pieces particle filtering half mask meet the requirements			
				The two samples both have a wider visual field				
9	Field of vision	—	The field of vision is acceptable if determined so in practical performance tests.	The two samples both have a wider visual field			Qualified	
10	Penetration of filter material	Sodium chloride	$\leq 6\%$	A.R.	0.1%	0.1%	0.1%	Qualified
				S.W.	0.2%	0.1%	0.2%	
		Paraffin oil	$\leq 6\%$	A.R.	0.6%	0.7%	0.6%	Qualified
				S.W.	0.7%	0.6%	0.7%	
				M.S+T.C.	1.8%	1.9%	1.8%	

Test Report

[2020] WSZ FHL NO.5420

Page 4 of 5

S.No	Test item	Unit	Technical requirements	Test result						Single item decision	
				Exercises	Facing directly ahead	Facing vertically upwards	Facing vertically downwards	Lying on the left side	Lying on the right side		
11	Breathing resistance	Inhalation 30 L/min	≤ 0.7	A.R.	0.3	0.4	0.3	0.4	0.3	Qualified	
					0.5	0.5	0.5	0.4	0.4		
					0.4	0.5	0.4	0.5	0.4		
				S.W.	0.4	0.4	0.4	0.5	0.5		
					0.5	0.4	0.5	0.4	0.5		
					0.4	0.5	0.5	0.4	0.4		
	T.C.	0.4	0.4	0.5	0.5	0.4					
		0.5	0.5	0.4	0.4	0.4					
		0.4	0.4	0.5	0.5	0.4					
	Breathing resistance	Inhalation 95 L/min	mbar	≤ 2.4	A.R.	2.1	2.2	2.1	2.2	2.2	Qualified
						2.2	2.1	2.2	2.1	2.1	
						2.2	2.1	2.1	2.2	2.2	
S.W.					2.1	2.2	2.2	2.1	2.1		
					2.2	2.2	2.1	2.1	2.2		
					2.2	2.2	2.2	2.1	2.1		
T.C.		2.2	2.1	2.2	2.1	2.2					
		2.2	2.2	2.2	2.1	2.1					
		2.2	2.2	2.2	2.2	2.1					
Exhalation 160 L/min		mbar	≤ 3.0	A.R.	2.7	2.8	2.7	2.8	2.7	Qualified	
					2.8	2.8	2.7	2.7	2.7		
					2.8	2.8	2.7	2.7	2.8		
	S.W.			2.8	2.8	2.7	2.7	2.8			
				2.8	2.7	2.7	2.8	2.8			
				2.7	2.8	2.8	2.7	2.7			
T.C.	2.8	2.8	2.7	2.7	2.7						
	2.7	2.8	2.7	2.8	2.7						
	2.8	2.8	2.7	2.7	2.8						

Test Report

[2020] WSZ FHL NO.5420

Page 5 of 5

S.No	Test item	Unit	Technical requirements	Test result							Single item decision	
				Exercises	E1 (%)	E2 (%)	E3 (%)	E4 (%)	E5 (%)	TIL (%)		
12	Total inward leakage	—	At least 46 out of the 50 individual exercise results shall be not greater than 11%; And in addition, at least 8 out of the 10 individual wearer arithmetic means for the total inward leakage shall be not greater than 8%.	A.R.	1 st	6.0	6.5	6.7	6.8	6.2	6.4	Qualified
					2 nd	6.2	6.9	7.3	7.1	6.6	6.8	
					3 rd	5.7	6.6	6.5	6.3	5.9	6.2	
					4 th	6.3	7.1	7.2	7.0	6.5	6.8	
					5 th	6.6	7.2	7.2	7.0	6.6	6.9	
				T.C.	6 th	6.5	6.8	7.2	7.2	6.5	6.8	
					7 th	5.9	6.3	6.4	6.7	6.0	6.3	
					8 th	6.3	6.8	6.9	7.2	6.5	6.7	
					9 th	6.6	7.7	7.6	7.6	7.0	7.3	
					10 th	6.8	7.8	7.6	7.5	7.1	7.4	
Note												

The end

